

Calibration Certificate

This is to certify that ADVIA 1800 Sr.No. CA1291013601360 installed at, Tata 1MG Lab, Dehradun has been duly calibrated on 06/01/2023 and the same was verified with the following tests.

OPTICS:

Lamp Energy Check:-

WL (NM)	A/D VALUE	VOLT	ATTENUATION %	REMARKS
340	10869	3.3	100	Verified
410	12734	3.8	100	Verified
451	13463	4.1	100	Verified
478	14665	4.4	100	Verified
505	15822	4.8	100	Verified
545	16364	4.9	100	Verified
571	16963	5.1	100	Verified
596	17549	5.3	100	Verified
658	18764	5.7	100	Verified
694	19249	5.8	100	Verified
751	20870	6.3	100	Verified
805	21440	6.5	100	Verified
845	22492	6.8	100	Verified
884	23974	7.3	100	Verified

Lamp Voltage 11.5 V (+/- 0.5 VDC) = 11.9 V

Verified

ELECTRICAL SYSTEM:

5 VDC Power Supply at LLS Connector (5VDC +/- 0.5VDC)

TEST POINT	VDC	REMARKS
DPP	5.01	Verified
RPP 1	5.02	Verified
RPP 2	5.03	Verified

24 VDC Power Supply at PMD & DC Driver2 (24VDC +/- 0.5VDC)

TEST POINT	PCB	VDC	REMARKS
PS1	PMD 14	24.01	Verified
PS2	PMD 24	24.0	Verified
PS3	PMD 34	24.03	Verified
PS4	DC DRIVER 2	24.02	Verified

HYDRAULICS

LWP Pressure (76 pKa)=76.5 KPa

VP Vacuum (-78 KPa to -91KPa)= -79 KPa

RRV ASSEMBLY

RRV Bath Oil Rate (4600 to 4800 ml/min)=4610

TEMPERATURES

TRAY	MEASURED TEMP	RANGE	REMARKS
CTT	6	6 to 14	Verified
RTT-1	7.2	6 to 14	Verified
RTT-2	7.1	6 to 14	Verified
RRV	37.0	37 +/-5	Verified

Verified the Calibrations and alignment of Robotics visually.

Calibration of the instrument was carried out as per the recommendation.

Next Calibration will be performed on 05.07.2023



Gajendra Shekhawat
Senior Executive-Customer Service

Service Report

Tata 1MG Technologies Pvt. Ltd.

3rd Floor, 1072, Ashirwad Tower
Dehradun 248001

Date: 06.01.2023

Ticket No: 728201299431

Service Type: Medical Maintenance

Complainant Name:

Complainant Phone:

The following services have been provided for Ticket No. 728201299431


Equipment Details			
Model:	ADVIA 1800		
Equipment Id:	1037656584	Serial No.:	CA1291013601360
Location:	664-150162		
Contract No:	2600461009	Contract Type:	MED Warranty

Service Call Reason
HDX India 6 Monthly visit/ 2PM

Observations & Actions Performed	
06.01.2023	Cleaned interior and exterior of instrument Performed jeol maintenance Performed daily maintenance Repaired Vacuum pump Performed light energy monitor, found ok Performed cell blank, found ok Run QC, found ok Preventive maintenance performed

Service Call Efforts					
Date	ID	Engineer's / Application's Name	Start	End	Hrs:Min
06.01.2023	65764583	SHEKHAWAT,GAJENDRA	16:30	18:00	01:30
06.01.2023	65764583	SHEKHAWAT,GAJENDRA	18:00	23:00	05:00

Material Consumed	
Description	QTY



teamplay Fleet

Accessible on all devices, anytime, anywhere!

Streamline your fleet management.
Optimize your asset performance.

Download the app today



Android



iOS

Service Report

to: 1872 Technologies Pvt Ltd

Date: 26-01-2025

to: 1872, Address: Taramani
Chennai - 600020

Case No: 120251200001

Service Type: Medical Management

Completion Date:

Completion Place:

The following services have been provided by Team No. 120251200001

Task Used:

120251200001

Manoj

SI - 1872
26-01-2025

Signature: [Handwritten Signature]

Signature: [Handwritten Signature]
26-01-2025



teamploy Fleet

Available on all devices, anytime, anywhere!

Streamline your fleet management
Optimize your asset performance.

Scan QR to access



1872

1872