



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL LAB CENTRIFUGE

Certificate No:	22900004556	Issued On:	13/10/2022
ULR No:	CC214622900004556F	Calibration Date:	11/10/2022
Job Identification No:	ICS/C/AS/10/4556	Next Calibration Date:	10/10/2023
Ref. No:	SRF, Dated - 08/10/2022		

CALIBRATED FOR:	M/s Tata 1mg Technologies Pvt. Ltd., 2 nd Floor, Plot No-1072, Ashirwad Tower, Ballupur Road, Chakrata Road, Sunder Vihar, Uttarakhand, Dehradun-248001
------------------------	---

EQUIPMENT DESCRIPTION			
Name	Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/ Model	Remi/ R-8C Plus	Visual Inspection	Ok
Serial No.	ZHIN-36002	I.D. No.	---
Range	Up to 6000 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS			
Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH

STANDARD REFERENCE DETAILS					
Traceable to National/ International Standard					
Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023
USED EQUIPMENT DETAILS		Not Applicable			
PRINCIPLE/METHODOLOGY OF CALIBRATION:		As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)			

RESULTS: Mechanical Calibration

MEASUREMENT: (in RPM)		
CENTRIFUGE READING	CORRECTED VALUE OF TACHOMETER READING (AVERAGE)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1240	1244.8	± 6
1920	1926.2	± 6
3350	3362.3	± 7
4620	4636.1	± 7
5850	5871.7	± 7

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by *Sundh*
 Authorised Signatory
[Signature]
 Operation-Head