



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL TIMER

Certificate No:	23500000787	Issued On:	08/02/2023
ULR No:	CC214623500000787F	Calibration Date:	06/02/2023
Job Identification No:	ICS/C/E-T/02/787	Next Calibration Date:	05/02/2024
Ref. No:	SRF, Dated - 03/02/2023		

<b>CALIBRATED FOR:</b>	<b>M/s Life Well Diagnostic Pvt. Ltd.</b> Ground Floor, M-16, Lajpat Nagar-2 New Delhi-110024
------------------------	---

### EQUIPMENT DESCRIPTION

Name	Digital Timer Fitted Centrifuge,		
Make/Model	Euro / Electra BL (12T)	Visual Inspection	Ok
Serial No.	202106223-12	I.D. No.	LC/PNS/LEQ/CUG/GZB/12-21/006
Range	Up to 99 min.	Least Count	1 min.
Location	Biochemistry	Calibration Site	On-Site

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

### STANDARD REFERENCE DETAILS Traceable to National/ International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Stop watch	Racer	23500000004	Indian Calibration Services, Delhi	02/01/2023	02/01/2024

### USED EQUIPMENT DETAILS

N/A

### PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-ET12 (Comparison Method)

### RESULTS: Electro Technical Calibration (Time)

U.U.C. TIME SET	STD. OBSERVED READING (AVERAGE)	Uncertainty at approx 95% C.L and coverage factor k=2
01 Min.	01 Min. 00.15 s	± 0.6 min.
05 Min.	05 Min. 00.26 s	± 0.6 min.
15 Min.	15 Min. 00.46 s	± 0.6 min.
30 Min.	30 Min. 01 s	± 0.6 min.

### REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*



- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
**Indian Calibration Services**

Checked by

Authorised Signatory

Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL STOP WATCH

Certificate No:	23500000004	Issued On:	03/01/2023
ULR No:	CC214623500000004F	Calibration Date:	02/01/2023
Job Identification No:	ICS/C/ET/01/004	Next Calibration Date:	02/01/2024
Ref. No:	SRF, Dated - 02/01/2023	Receipt Date:	02/01/2023

<b>CALIBRATED FOR:</b>	<b>M/s INDIAN CALIBRATION SERVICES,</b> 313 & 304, Laxmi Deep District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi - 110092
------------------------	--

### EQUIPMENT DESCRIPTION

Name	Digital Stop Watch		
Make/Model	Racer	Visual Inspection	Ok
Serial No.	---	I.D. No.	ICS/ET/INS/05
Range	Upto 3600 Sec.	Least Count	0.01/ 1 Sec.
Location	---	Calibration Site	In-Lab

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Time Calibrator	PNJCS	C-220907-3-1	Godrej	07-12/09/2022	12/09/2023

USED EQUIPMENT DETAILS	N.A
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-ET12 (Direct Method)

### RESULTS: Electro-Technical Calibration (Time)

STD. READING	U.U.C. OBSERVED READING (AVERAGE)	Error	Uncertainty at approx 95% C.L and coverage factor k=2
1.000 s	00 Hr. 00 Min. 01.05 s	0.05 s	± 0.04 s
60.000 s	00 Hr. 01 Min. 00.15 s	0.15 s	± 0.12 s
900.000 s	00 Hr. 15 Min. 00.27 s	0.27 s	± 0.12 s
1800.000 s	00 Hr. 29 Min. 59.45 s	-0.55 s	± 0.24 s
3600.000 s	00 Hr. 59 Min. 58 s	-02 s	± 0.75 s

### REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.
- \*\*\* END OF REPORT \*\*\***

- Note: 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Satish  
Checked by

Authorised Signatory

RAJESH KAPOOR  
Operation-11-13