



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROPIPETTE

Certificate No:	23700003925	Issued On:	02/04/2023
ULR No:	CC214623700003925F	Calibration Date:	31/03/2023
Job Identification No:	ICS/C/MVD/03/3925	Next Calibration Date:	30/03/2024
Ref. No:	SRF, Dated - 30/03/2023	Receipt Date:	31/03/2023

CALIBRATED FOR: M/s Life Well Diagnostic Pvt. Ltd.
 Ground Floor, M-16, Lajpat Nagar-2
 New Delhi-110024

EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	---	Visual Inspection	Ok
Serial No.	YEAA9ADO012887	I.D. No.	LC/PNS/LEQ/PIPT/GZB/12-21/001
Range	100 µl - 1000 µl	Least Count	5 µl
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Digital Weighing Balance	AND	23700000002	Indian Calibration Services, Delhi	03/01/2023	03/01/2024

USED EQUIPMENT DETAILS N/A
PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

RESULTS: Mechanical Calibration (VOLUME)

Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.100 (0100 µl)	0.09991 (0099.91 µl)	± 0.83 µl
0.500 (0500 µl)	0.49985 (0499.85 µl)	± 0.83 µl
1.000 (1000 µl)	0.99974 (0999.74 µl)	± 0.83 µl

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***



For and on behalf of Indian Calibration Services

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

Checked by **RAJESH KAPOOR**
 Authorised Signatory
 Operation Head



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 2

CALIBRATION CERTIFICATE OF DIGITAL WEIGHING BALANCE

Certificate No:	23700000002	Issued On:	04/01/2023
ULR No:	CC214623700000002F	Calibration Date:	03/01/2023
Job Identification No:	ICS/C/MVD/01/002	Next Calibration Date:	03/01/2024
Ref. No:	SRF, Dated - 02/01/2023		

CALIBRATED FOR:
M/s Indian Calibration Services.
304 & 313, Laxmi Deep District Centre.
Near Nirman Vihar Metro Station.
Laxmi Nagar, Delhi - 110092

EQUIPMENT DESCRIPTION

Name	Digital Weighing Balance		
Make/Model	AND/ BM-252	Visual Inspection	Ok
Serial No.	T1002614	I.D. No.	ICS/MVD/INS/02
Range	Up to 250 g	Least Count	0.00001 g
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS	Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
--------------------------	-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Weights	Weightronics	SCW-1222-405	Maruti, Ahmedabad	31/12/2022	30/12/2025
USED EQUIPMENT DETAILS		N.A.			
METHODOLOGY OF CALIBRATION		As per Calibration Procedure No.: ICS/CAL/SOP-M02			

RESULTS: Mechanical Calibration (Mass)

1. Linearity Test

Sr No.	Nominal Mass Value	Mass Value Of Standard (g)	Balance Reading (g) (Avg.)
1.	0	0	0.00000
2.	1 mg	0.0009993	0.00100
3.	10 mg	0.0099994	0.01000
4.	50 mg	0.0499983	0.05000
5.	100 mg	0.1000009	0.10000
6.	500 mg	0.4999970	0.50000
7.	10 g	10.000011	10.00001
8.	50 g	50.000008	50.00001
9.	100 g	100.000014	100.00002
10.	200 g	200.00001	200.00002
11.	250 g	250.000018	250.00003

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory

Operator