



PT EQAS RESULT REVIEW FORM

Issued by QA:

Department: <i>Clinical Pathology</i>	Month / Year: <i>MAY-2023</i>
PT Provider: <i>METROPOLIS MHL EQAS</i>	Distribution / Lot No: <i>2304/01</i>
Analyte(Test): <i>Crystal.</i>	
Machine used: <i>MICROSCOPIC</i>	Kit used: <i>microscopic.</i>
Test done by: <i>DR. SHWETA AHUJA</i>	Done on: <i>12/5/2023</i>
Report uploaded by: <i>AMIT KUMAR</i>	Report uploaded on: <i>14/5/2023</i>
Software used: <i>METROPOLIS MHL EQAS</i>	
EQAS Report received on <i>19/5/2023</i>	
Reviewed by: <i>DR. SHWETA AHUJA</i>	Approved by: <i>DR. SHWETA AHUJA</i>
Observations: <i>ALL EQAS RESULT IS Satisfactory with 100.0% score.</i>	

*If results are ungraded, please attach self-assessment. If satisfactory no corrective and preventive action required. But if unsatisfactory, corrective and preventive action is required with optimum evidence.

Self-Assessment Result: Satisfactory Unsatisfactory

Attachment: Yes No

Please investigate failure for unsatisfactory performance and submit a report to QA and Medical Director in 3 Days.

Approved by *DR. SHWETA AHUJA*

Signature
(Lab Director/Designee)



Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry			

Problem Classification: _____
 _____ *NA*

Corrective Action: _____
 _____ *NA*

Reviewed by: *Kuldeep Singh*
Assistant manager
Chandigarh Lab



Dated: 7/6/2023

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Struvite crystals	Struvite crystals
Legend @ : Acceptable		Total Parameters	1
⊕ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	1
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070



PT EQAS RESULT REVIEW FORM

Issued by QA:

Department: <i>Clinical Pathology</i>	Month / Year: <i>May-2023</i>
PT Provider: <i>METROPOLIS MHL EQAS</i>	Distribution / Lot No: <i>230104/01</i>
Analyte(Test): <i>pH, Specific gravity, Glucose (urine), Protein (urine), Ketone (urine), Blood (urine), Urobilinogen (urine), Nitrate (urine).</i>	
Machine used: <i>ABBOTT UROMETER-120</i>	Kit used: <i>ABBOTT Urometer urine strip</i>
Test done by: <i>Ritika</i>	Done on: <i>12/5/2023</i>
Report uploaded by: <i>AMIT KUMAR</i>	Report uploaded on: <i>14/5/2023</i>
Software used: <i>METRO POLIS MHL EQAS</i>	
EQAS Report received on <i>19/5/2023</i>	
Reviewed by: <i>AMIT KUMAR/KULDEEP</i>	Approved by: <i>DR. SHWETA AHUJA</i>
Observations: <i>ALL EQAS RESULT within Acceptable Range with 100.0% Score EQAS Result is satisfactory</i>	

*If results are ungraded, please attach self-assessment. If satisfactory no corrective and preventive action required. But if unsatisfactory, corrective and preventive action is required with optimum evidence.

Self-Assessment Result: Satisfactory Unsatisfactory

Attachment: Yes No

Please investigate failure for unsatisfactory performance and submit a report to QA and Medical Director in 3 Days.

Approved by *DR. SHWETA AHUJA*

Signature
(Lab Director/Designee)



Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable			
X : Unacceptable			
# : Not Evaluated			
⊙ Delayed Result Entry			


Problem Classification: _____

 _____ *NA*

Corrective Action: _____

 _____ *NA*

Reviewed by: *Kuldeep*
7/6/2023
Assistant manager
Chandigarh - Lab



Dated: *7/6/2023*

Instrument : Urometer 120

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5,5,5,6
@ Specific Gravity	-	1.025	1.015,1.020,1.025
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	Trace	+1,Tr
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	Negative	Neg
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
: Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	0
⓪ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070