



Rapid Diagnostic Pvt. Ltd.



ISO 9001:2015
CERTIFIED COMPANY

CALIBRATION CERTIFICATE

CUSTOMER NAME : Sri Vishnu Health Care Unit,
(A Unit of Tesla Diagnostics)
D.No:10-35/4, Rajesh Towers, Pothireddy Pally,
Sanga Reddy Dist, T.S

INSTRUMENT NAME : Fully Automated Biochemistry Analyzer.

MODEL : CHEM ULTRA.

SERIAL No : AES1J9WH0418Y.

THIS INSTRUMENT IS CALIBRATED ON DATE: 10/04/2023 IS VERIFIED CALIBRATION / WORKING
CONDITION ACCORDING TO INSTRUCTED BY THE MANUFACTURER. THE SYSTEM PARAMETERS ARE
FOUND TO BE WITHIN THE SPECIFIED RANGES. THE SYSTEM HAS PASSED THE PERFORMANCE
TESTS. THE INSTRUMENT IS CALIBRATED WITH THE RESPECTIVE CALIBRATORS.

CALLIBRATION VALIDITY : 12 Months.

VERIFIED BY : N. Mahender Reddy (Regional Tech. Manager)


CALIBRATED DATE : 10/04/2023

NEXT CALIBRATION DATE : 09/04/2024

For RAPID DIAGNOSTIC PVT. LTD

AUTHORIZED SIGNATORY.

N.M. Reddy



RAPID DIAGNOSTIC PVT. LTD.
Plot No.22, Lane No.01, Vahini Nagar Colony,
Opp:OBC Bank Lane, Sikh Road,
Bowenpally, Secunderabad-500 009.

B-82, G.T. Karnal Road Industrial Area, Delhi-110 033, Ph.: 91-11-45201100, Fax: 91-11-27413437, 45201198
e-mail: drjindal@rdgc.in, website: <http://www.rapiddiagnostic.in>



First Line

Second Line

QC(Name-Lot)	Test	Target Value	SD Value	Result	Test Date	State
SERA	GIUCOSE	92.00	9.00	87.53	10-04-2023 08:33:43	
SERA	CREATIN	1.23	0.15	1.44	10-04-2023 08:34:07	
SERA	UREA	40.00	3.20	36.17	10-04-2023 08:34:31	
SERA	CHOL	127.00	8.50	143.16	10-04-2023 08:35:19	
SERA	TG	117.00	10.00	109.78	10-04-2023 08:35:43	
SERA	D.BILL	0.62	0.10	0.68	10-04-2023 08:36:31	
SERA	SGOT	38.40	4.50	35.42	10-04-2023 08:36:55	
SERA	SGPT	37.30	4.50	32.90	10-04-2023 08:37:19	
SERA	ALP	78.10	7.00	77.45	10-04-2023 08:37:43	
SERA	ALB	3.20	0.40	3.85	10-04-2023 08:38:31	
SERA	HDL	37.00	4.00	41.32	10-04-2023 08:38:55	
SERA	CAL	8.50	0.40	9.07	10-04-2023 08:39:19	