



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



Calibration Certificate

Format No.: UCCS/QF/7.8-05

Page No :- 1 of 1

Certificate No.	UCCS/21/A/167-009	<h3>Calibration Certificate of</h3> <h3>Micro pipette</h3>
Issue Date	11.03.2023	
Service Request No.	UCCS//21/H/166	
Service Request Date	11.03.2023	
Receipt Date	11.03.2023	
Calibration Date	11.03.2023	
Calibration Due Date (As per Customer request)	10.03.2024	

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
DR SUNITA HOSPITAL Js Nagar Nawabganj Kaimganj Farrukhabad District Uttar Pradesh	Range/ Size Make/Model Instrument I.D. No. Instrument Sr. No. Least Count Location/Dept. Visual Inspection	100 to 1000 ul Perfect Xtra/REF-T/ML/11 0.1 ul MAIN LAB Ok

Calibration Procedure	UCCS/TH/CP- 04	Condition of Instrument	OK
Reference Standard	DKD R5-1	Calibration Performed at	Lab
Environmental Condition	Temperature	30 ± 5 °C	Relative Humidity 30 to 80 % RH

Standard Equipment Used for Calibration					
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.	Due Date
Tycon Engineering	TIPL	UCCL/ET/PC-01	Acumen Measurement Delhi	AMC/00001537	6/30/2024

Serial No	Nominal Value In (kg/cm ²)	Observed Value In kg/cm ²	Error in °C Mean	Uncertainty of Measurement ± °C
1	200.0	199.9	0.1	± 0.46 °C
2	500.0	499.8	0.2	
3	1000.0	999.7	0.4	

The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %

Note :-

1. This Certificate refers only to the particular items submitted for calibration.
2. This Certificate shall not be reproduced, except in full, without the written permission of Managing director Calibration System, Rudrapur.
3. Result Reported are Valid at the of and under the stated conditions of measurement.
4. Laboratory Standards are traceable to National Standards.
5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment,



Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commercial use.

End of Certificate

Amit
Calibrated By
Amit Kumar
Calibration Engineer

Authorized Signatory
Vikash
Reviewed/Approved By
Vikash Kumar
Technical Manager

Working : W.Z. -1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046
Regd. Office : R.Z.-56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045