

DECEMBER - 2022

53



## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name Govt. PHC, T.Pottakollai Lab No 12010  
Constituent Group Chemistry III Date of Result Entered : 21/12/2022  
PT Item Lyophilized human serum based Date of Report Published : 05/01/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1844	239.20	6.25	14.96	298 mg/dL	3.93	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1839	3.91	8.24	0.32	3.8 mg/dL	-0.34	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1766	114.94	5.41	6.21	115 mg/dL	0.01	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### LAB ADDRESS :

Govt. PHC, T.Pottakollai  
THE MEDICAL OFFICER, T.POTTAKOLLAI  
T.Pottakollai  
TAMILNADU621804

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

மருத்துவ அலுவலர்  
அக அரம்ப கலாதர நிலையம்,  
த.மொட்டக்கொல்லம்,  
அரியலூர் மாவட்டம் - 621 804

