

VIEW LAB MONTHLY SUMMARY

Lab Name	SURAHA SPECIALITY LABORATORY		Details About Robust Analysis							
Lab No	16610		Detail About Monthly Summary							
Month	February		Detail about SDI							
Year	2023									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 18/02/2023							
			Date of Report Published : 08/03/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	494	119.76	7.78	9.32	126.5 mg/dL	0.72	0.84
2	UREA II	Urease UV / GLDH II	Agape	413	125.79	9.60	12.08	149.2 mg/dL	1.94	1.19
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	358	4.19	9.72	0.41	4.64 mg/dL	1.11	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	446	1.58	21.85	0.34	1.87 mg/dL	0.84	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	314	4.68	8.32	0.39	4.73 g/dL	0.13	0.04
6	ALBUMIN II	BCG - Colorimetric II	Agape	335	2.99	10.41	0.31	3.49 g/dL	1.61	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	384	7.16	10.39	0.74	14.9 mg/dL	10.46	0.08
8	CHOLESTROL II	CHOD-PAP II	Agape	488	96.79	12.52	12.11	104.8 mg/dL	0.66	1.10
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	328	84.47	11.21	9.47	86.1 mg/dL	0.17	1.05
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

