

VIEW LAB MONTHLY SUMMARY

Lab Name	SURAHA SPECIALITY LABORATORY	Details About Robust Analysis									
Lab No	16610	Detail About Monthly Summary									
Month	April	Detail about SDI									
Year	2023										
Constituent Group	Chemistry II										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 18/04/2023									
		Date of Report Published : 06/05/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE II	GOD-POD II	Agape	524	149.18	8.49	12.67	195.4 mg/dL	3.65	1.11	
2	UREA II	Urease UV / GLDH II	Agape	479	37.53	11.09	4.16	46.7 mg/dL	2.20	0.38	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	394	2.48	9.40	0.23	2.61 mg/dL	0.56	0.02	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	480	3.12	16.43	0.51	4.49 mg/dL	2.68	0.05	
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	371	5.54	8.14	0.45	6.69 g/dL	2.55	0.05	
6	ALBUMIN II	BCG - Colorimetric II	Agape	367	3.38	9.46	0.32	3.15 g/dL	-0.72	0.03	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	449	5.54	14.33	0.79	8.78 mg/dL	4.1	0.07	
8	CHOLESTROL II	CHOD-PAP II	Agape	523	117.23	11.15	13.08	135.2 mg/dL	1.37	1.14	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	386	103.22	12.42	12.82	131.6 mg/dL	2.21	1.30	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

