



CALIBRATION CERTIFICATE

ADL/DO/SER/0024/23-24

Customer Name & Address	Shifa Hospital, C/o Simple Medical Solutions, 59, Pattukkottai Road, Thanjavur, Tamil Nadu, Adirampattinam, Tamil Nadu-614701, India
-------------------------	--

INSTRUMENT	INSTALLED ON	CALIBRATED ON
Mispa I3	15-11-2019	29/Mar/2023

EQUIPMENT DETAILS

Instrument name	Mispa I3
Serial No	21192210542
Environmental condition:	Temperature:
Equipment calibration	
Parameter calibration	Inhouse (Calibrated IC card provided along with each kit)

This is to certify that the above-mentioned product has been calibrated & tested for the performance of kits with Control.

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaishingh Business Centre, 119,
Sahar Road, Parsipada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



Equipment Test report

Checked the IC Card Reader	: Found Ok
Checked the Probe	: Found Ok
Checked Mother Board	: Found Ok
Checked the Wash Assembly	: Found Ok
Channel-1 (with cartridge)	: 2,67,227
Channel-1 (without cartridge)	: 3,27,162
Channel-2 (with cartridge)	: 2644
Channel-2 (without cartridge)	: 332

Performance validation with Control

HbA_{1c}
Protein Control

LOT NO: 33010072 EXP: 10/2024.

Parameter	Low	High	Result	Target
GRP HbA _{1c}	4.6	6.2	5.6	5.4
RF				
ASO				

Next Calibration due on 28/Sept/2023

For Agappe Diagnostics Ltd,



Authorized Signatory

Name: H.Sriram

Manager- Asst. Manager - Customer service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PQ), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector-V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsipada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5/6, IInd Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in