



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name **PATHOLOGY ,BETUL** Lab No **7789**
Constituent Group **Chemistry I** Date of Result Entered : **15/04/2023**
PT item **Lyophilized human serum based** Date of Report Published : **08/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Biosystems	325	150.75	6.88	10.38	159.07 mg/dL	0.80	1.15
2	UREA	Urease UV / GLDH	Biosystems	281	35.88	8.65	3.10	32.46 mg/dL	-1.10	0.37
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Biosystems	237	2.85	8.63	0.25	2.95 mg/dL	0.41	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Biosystems	247	2.69	16.93	0.46	2.15 mg/dL	-1.19	0.06
5	T-PROTEIN	Biuret - Colorimetric	Biosystems	300	5.55	5.95	0.33	5.6 g/dL	0.15	0.04
6	ALBUMIN	BCG - colorimetric	Biosystems	301	3.27	6.82	0.22	3.14 g/dL	-0.58	0.03
7	CALCIUM	Arsenazo III	Biosystems	201	11.38	8.26	0.94	12.06 mg/dL	0.72	0.13
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Biosystems	92	3.65	13.62	0.50	3.8 mg/dL	0.30	0.10
9	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	297	6.09	8.83	0.54	6.51 mg/dL	0.78	0.06
10	CHOLESTEROL	CHOD-PAP	Biosystems	299	118.99	6.82	8.12	112.48 mg/dL	-0.80	0.94
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	283	105.47	9.35	9.87	109.33 mg/dL	0.39	1.17
12	HDL	Direct method / Enzymatic colorimetric	Biosystems	206	27.87	14.03	3.91	25.08 mg/dL	-0.71	0.55
13	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1181	131.83	3.23	4.26	126.18 mmol/L	-1.33	0.25
14	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1175	5.43	5.95	0.32	5.37 mmol/L	-0.19	0.02
15	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	1011	102.67	3.93	4.03	101.11 mmol/L	-0.39	0.25
16	AST	UV kinetic(with & without PLP (P-5-P))	Biosystems	296	103.61	11.15	11.55	105.51 U/L	0.16	1.34
17	ALT	UV kinetic(with & without PLP (P-5-P))	Biosystems	286	77.12	13.46	10.38	100.19 U/L	2.22	1.23
18	ALP	PNP AMP kinetic	Biosystems	294	108.54	15.05	16.33	148.95 U/L	2.47	1.90
19	AMYLASE	Enzymatic Colorimetric / G7PNP Blocked	Any Analyser (Automation / Semi Automation)	257	45.47	20.28	9.22	44.65 U/L	-0.09	1.15
20	IRON	Ferro Zine (No Protein Removal)	Biosystems	61	155.46	9.19	14.28	152.09 ug/dL	-0.24	3.66

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
PATHOLOGY ,BETUL
DISTRICT HOSPITAL BETUL
BETUL
MADHYA PRADESH460001

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******* End of Report *******