



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :-	Page No.	:- 1 of 1
ICTC - 2 CENTRE	SRF No	:- GTS/220927/02
North Goa District Hospital, MAPUSA, GOA	Certificate No.	:- GTS/220927/02-001
	Date of Received	:- 27.09.2022
	Date of Calibration	:- 27.09.2022
	Next Calibration Due On	:- 26.09.2023
	Issue Date	:- 03.10.2022
Ambient Temp. (°C) :- 21 to 40	Calibration method No.	:- MECH-WI-07
Relative Humidity (%RH) :- 30 to 75	ULR No	:- CC295722000010124F
Location of calibration :- On Site		
Condition of Item :- Ok		

2. Description of Item

Name :- Centrifuge	Range :- 0 to 3500 RPM
Id No :- DHS/MSD/PATH/B2	Least Count :- -
Make :- Remi R303	Location :- ICTC 2 LAB
Type :- Knob	
	Sr No :- FHLC 10191
	Dept. :- GSACS

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Digital Tachometer	NI/2205/059/003	Nishitronics Instrumentation	GTS/TM/01	22.05.2023

* Discipline-Mechanical

4.Calibration Results :-

Sr No	Standard Reading RPM	Knob No on UUC	Expanded Uncertainty in ± %
1	859.7	1	4.20
2	1452	2	1.20
3	2411	3	1.20
4	2818	4	1.20
5	3419	5	1.20

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6)The Instruments used for calibration are traceable to National/International standards and their Calibrations are valid.

Calibrated By

Rohit
Calibration Engineer
Rohit L



Approved By

SB
Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate



CC-2294



Nishitronics



Calibration Services

Pune-411044

Phone: 020-26111111

Fax: 020-26111112

Email: nishitronics@gmail.com

CALIBRATION CERTIFICATE					
CUSTOMER :- Global Technical Services Sec. No-25, Plot No-49/3 LIG Colony Pradhikaran Nigdi Pune-411044		Page No. :- 1 of 1 ULR No. :- ULR-CC229422000002373F Discipline :- Mechanical Calibration Certificate No :- NI/2205/059/003 Date of issue :- 24/05/2022 Date of receipt :- 23/05/2022 Date of calibration :- 23/05/2022 Cal. Req. No. :- NI/2205/059/003 Next Due Date :- 22/05/2023 Parameter :- RPM Calibration method no :- NI / CP / M / 05			
Amb. Temp :- 25 ± 2 °C Rh. :- 45 to 75% Location of calibration :- LAB Characteristic and :- OK Condition of items :- OK					
Details of Items					
Name :- Digital Tachometer Location :- -- Make :- Metrix ID No. :- GTS/TM/01 Sr No. :- N785275	Model :- TM-4005 Range :- 2.5 To99999 rpm(non-contct) L.C. :- 0.1 rpm & 1 rpm Accuracy :- Not specified				
Details of Equipment used for Calibration					
Name :- DIG.TACHOMETER Calibrated by :- INSTOTECH Certificate No. :- 2021/0200 Validity :- 01/06/2022					
OBSERVATION					
NON-Contact Type					
Cal. Point rpm	Std. Reading rpm	UUC Reading rpm	Error rpm	Expanded Uncertainty	
100	100.1	100.6	0.500	± 3.0 rpm.	
5000	5001	5002	1.000	± 5.0 rpm.	
10000	10115	10109	-6.000	± 10.0 rpm.	
20000	20548	20538	-10.000	± 15.0 rpm.	
40000	42350	42342	-8.000	± 20.0 rpm.	
90000	90326	90318	-8.000	± 25.0 rpm.	
The reported measurement uncertainty is estimated at a level of confidence of approximately 95 %					
Remarks :- 1) Result are related only to the item calibrated . 2) This certificate refers only to the particular items submitted for calibration . 3) This certificate shall not be reproduced except in full without our prior permission in writing. 4) The calibration results reported in this particular certificate are valid at the time of an under stated condition of measurement. 5) Standard used for calibration were traceable to National / International standard. 6) Readings given above are as on received condition of an instrument.					
Calibrated by  (P.B. Desai) (Technical Assistant)		Authorized Signatory  (V.B.Hingmire) / (S.B.Hingmire) (Technical Manager)			
Form No : NI/F/7.8/ET/01		Form Rev. No. : 0		Effective Form Date:20/11/2019	
Issue No : 03		Issue Date :20/11/2019			