

Erba Lachema s.r.o.




Brno, 10.11.2016

To,
Redcliffe Lifetech Pvt Ltd.
Jabalpur, Madhya Pradesh.

Instrument Name: ERBA LAURA
Instrument Serial no: 1003-1-202206

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

P. F.

RNDr. Pavel Suchan Ph.D.

Erba Lachema s.r.o.
Karásek 1d, 621 00 Brno
DIČ: CZ26918846 -10-

Global Product Manager, Urine Analysis
Erba Lachema s.r.o.
Karásek 1d, Brno 62100, Czech Republic



DEKAPHAN LAURA
Seq.No: 0002
URINORM P
2023.06.20 05:57

.....
*BLD 250 Ery/u1 3+
*BIL 6 mg/dl 3+
*UBG 3 mg/dl 2+
*KET 52 mg/dl 2+
*GLU 500 mg/dl 2+
*PRO 500 mg/dl 3+
PH 9
*NIT POS
*LEU 500 Leu/u1 3+
SG 1.000

QC TEST: OK
<Admin>

DEKAPHAN LAURA
Seq.No: 0001
URINORM N
2023.06.20 05:54

.....
BLD NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 6
NIT NEG
LEU NEG
SG 1.025



QC TEST: OK
<Admin>