


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221209A/N-117	Page	: 1 of 1	
CALIBRATED ON.	: 10 / 12 / 2022	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 09 / 12 / 2023	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 221209A/N Dt. 09.12.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000001082F Dt. 12.12.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. New-639, Old-295, Napier, First Floor Town Opposite Bhanwartal Garden, Jabalpur			
ITEM CALIBRATED	: Digital Temp. Indicator			
	Make : RoHS	Range	: (-50 to 110) °C	
	L. Count :	ID No.	: RCL/SR/D.Temp-05	
	Location : Sample Receiving Dept.	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023
RESULTS	: ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.04	2	0.04		
4.06	4	0.06		
6.10	6	0.10		
8.12	8	0.12		
(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

***U.C.** : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida