


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221209A/N-108	Page	: 1 of 1	
CALIBRATED ON.	: 09 / 12 / 2022	Ref. Std.	: DKD R-5-7	
CALIBRATION DUE.	: 08 / 12 / 2023	C.P. No.	: ACS/CP/11	
SRF No. & DATE	: 221209A/N Dt. 09.12.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000001073F Dt. 12.12.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. New-639, Old-295, Napier, First Floor Town Opposite Bhanwartal Garden, Jabalpur			
ITEM CALIBRATED	: Digital Thermo Hygrometer			
	Make	: Digital Series	Range	: (0 to 50)°C, (10-99)%
	L. Count	:	ID No.	: RCL/BIO/HGM-02
	Location	: Clinical Biochemistry	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp. { 25°C + 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023
RESULTS	: Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR		
(°C)	15.04	14.99	0.05	
	25.06	24.98	0.08	
	40.06	39.97	0.09	
(%)	30.04	29.99	0.05	
	60.04	59.98	0.06	
	90.06	89.97	0.09	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.8^\circ\text{C}$, 1.9% at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

