



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name **FEROKE NEETHI MEDICAL LAB** Lab No **14870**
Constituent Group **Chemistry I** Date of Result Entered : **11/05/2023**
PT item **Lyophilized human serum based** Date of Report Published : **05/06/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Transasia / Erba	957	256.55	5.99	15.38	259 mg/dL	0.16	0.99
2	UREA	Urease UV / GLDH	Transasia / Erba	938	78.15	7.38	5.77	76 mg/dL	-0.37	0.38
3	CREATININE	Enzymatic Colorimetric	Transasia / Erba	483	1.55	7.15	0.11	1.35 mg/dL	-1.80	0.01
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Transasia / Erba	973	4.04	9.87	0.40	4.43 mg/dL	0.98	0.03
5	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	950	5.35	7.84	0.42	5.47 g/dL	0.29	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	985	3.17	5.51	0.18	3.32 g/dL	0.86	0.01
7	CALCIUM	Arsenazo III	Transasia / Erba	710	9.56	8.81	0.84	9.6 mg/dL	0.05	0.06
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Transasia / Erba	453	5.67	8.85	0.50	5.86 mg/dL	0.38	0.05
9	URIC ACID	Enzymatic / Uricase Colorimetric	Transasia / Erba	910	8.77	7.20	0.63	9.7 mg/dL	1.47	0.04
10	CHOLESTEROL	CHOD-PAP	Transasia / Erba	965	104.58	7.53	7.88	99 mg/dL	-0.71	0.51
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	924	115.34	8.86	10.22	111 mg/dL	-0.42	0.67
12	HDL	Direct method / Enzymatic colorimetric	Transasia / Erba	731	24.66	10.89	2.68	20 mg/dL	-1.74	0.20
13	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1209	134.44	3.09	4.16	130 mmol/L	-1.07	0.24
14	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1220	3.12	5.73	0.18	3.04 mmol/L	-0.45	0.01
15	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	1011	98.34	4.20	4.14	95 mmol/L	-0.81	0.26
16	AST	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	906	139.68	10.13	14.15	154 U/L	1.01	0.94
17	ALT	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	907	76.70	15.69	12.04	83 U/L	0.52	0.80
18	ALP	PNP AMP kinetic	Transasia / Erba	878	299.55	15.52	46.49	334 U/L	0.74	3.14

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

FEROKE NEETHI MEDICAL LAB

Run by Feroke Service Co-Operative Bank Ltd, NEAR ESI HOSPITAL, FEROKE

KOZHIKODE
KERALA673631

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******* End of Report *******