

TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED "Perfecting the art calibration in Medical Equipment's"

Contact Decreases (Contact Contact Con	CALIBRATIO	N / TESTIN	G REPORT	
CUSTOMER DETAILS: M/s. SRI ARUNA COMPUTER DIAGNOSTIC CENTRE, No:657/3, NH-45, Trichy Main Road, Veppur-606304 CALIBRATION DATE 04/05/2023			ICATE NO	TICS/23/05/017
				SEMI AUTO ANALYZER
			ACTURE	QUICKLAB
				QUICKCHEM PRO PLUS
			NO	EC90042231050
CALIBRATION DUE DATE	03/05/2024		LED DEPARTME	
1. General maintenance re	equired	☐ Yes	☑ No	
2. Any Physical damage	observed	☐ Yes	☑ No	
3. Spare Replacement (if	required)	☐ Yes	☑ No	
				1
TEST EQUIPMENT	MANUFACTURE	МО	1	RIAL NUMBER &
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini		M 18040011 & ACCS221441 -07-2023)
*The above mentioned instrument traceability to National/Internation		ards manufacture	s recommended proto	ocols, using equipments having
✓ ACCEPTABLE FOR USE ☐ CORRECTIVE MAINTENANCE REQUIRED				REMOVED FROM USE
Calibrated By:			Aut	thorized Sign With Seal
	AMOORTHY			TICS TO
*Electrical safety test result at				Callionna A OLY

PAGE:10f2



Resistance Point-to-Point

Resistance

Equipment

Applied Part

Leakage

Leakage

TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED

"Perfecting the art of calibration in Medical Equipment's"

TESTING REPORT							
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18	VPM 18040011			
TEST REPORT -Electrical Safety Test Standard: IEC 62353							
Parameter	Units	Value	Limits	Status			
Main Voltage L1-L2	V	231.4	240±10	PASS			
Main Voltage L1-G(PE)	V	232.0	240±10	PASS			
Main Voltage L2-G(PE)	V	3.1	<5	PASS			
Load Current	A	2.45	As per unit	PASS			
Protective Earth	0	0	<0.3	PASS			

Ω

Ω

μA

μA

The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

0

2.6

0



PASS

PASS

PASS

< 0.3

< 0.5

Class I - < 500

Class II - < 100

Type BF - < 5mA

Type CF- <50µA*

^{*} Note: Defibrillator paddles 100 µA, per IEC 60601-2-4