

NAVASHVAR CALIBRATION SERVICES

No. 303, Anand Building, Room No: 09, First Floor,
Poonamallee High Road, Kilpauk, Chennai -600 010.
Ph : +91-9087221939, 9791503514.

NAVASHVAR

E-mail: navashvar.info@gmail.com, Website : www.navashvar.com



CALIBRATION CERTIFICATE

Page 01 of 01

ULR Number: CC31452300001596F

Customer Name and Address: M/s. ARUNA DIAGNOSTIC CENTRE No: 657/3 NH.45, Trichy Main Road, Veppur-606304.	Certificate No	: CC-23/24-S-011/04
	SRF No	: SRF-23/24-S-011
	Date of Receipt	: 21/04/2023
	Date of Calibration	: 21/04/2023
	Due Date of Calibration	: 20/04/2024
	Date of Issue	: 24/04/2023

Details of Unit Under Calibration

Description	: Refrigerator	Accuracy	: --
Make	: LG	Asset ID.No	: NA
Model/ Type	: GL-B191KRLQ	Condition of UUC	: Good
Serial No.	: SADC-004	Calibrated At	: Site
Range	: 2 to 8 °C	Location	: NA
Resolution	: --		

Details of Reference Standard

Sl.No.	Description	ID.No/Sr.No	Certificate Number	Validity
1	Four wire RTD Sensor	NCS/Thermal/RTD02	TSC/22-23/11472-1	05/11/2023
2	Digital Thermometer	NCS/Thermal/DTM02	TSC/22-23/7495-1	10/08/2023

Environmental Condition	Temperature (°C)	25 ± 5	Reference Standard	DKD-R-5-7
	Humidity (%RH)	55 ± 20	Calibration Procedure	NCS-L03-WI-THER-005
Calibrated By	A. Naveenkumar			

Calibration Results

Results of Calibration in °C				
Sl.No.	Standard Reading (Mean)	UUC Reading (Mean)	Error	Measurement Uncertainty (±)
1	5.250	5	-0.250	0.63

Note:

1. The Standards used are traceable to National / International Standards.
2. This report refers only to the item calibrated and shall not be reproduced in full without the written approval of the issuing laboratory.
3. This report is not valid unless it is signed by the authorized signatory.
4. Next Calibration due date mentioned as per customer requirement.
5. The result is valid at the time of and under the stated conditions of measurement.
6. The measurement confidence level is 95.45% with coverage factor K=2.
7. Temperature Scale: ITS-90
8. UUC stands for Unit Under Calibration.



Authorised Signatory

A. Saranraj

A. Saranraj

Technical Manager

Document Number: NCS-L04-R42
Issue No. & Date: 01 & 19.08.2020