NAVASHVAR CALIBRATION SERVICES



No. 303, Anand Building, Room No: 09, First Floor, Poonamallee High Road, Kilpauk, Chennai -600 010.

Ph: +91-9087221939, 9791503514.



NATIONAL E-mail: navashvar.info@gmail.com, Website: www.navashvar.com

CALIBRATION CERTIFICATE	Page 01 of 01
The state of the s	

: 24/04/2023

		ULR Number: CC314523000001596F
Customer Name and Address:	Certificate No	: CC-23/24-S-011/04
M/s. SRI ARUNA COMPUTER DIAGNOSTIC CENTRE	SRF No	: SRF-23/24-S-011
No: 657/3 NH.45, Trichy Main Road,	Date of Receipt	: 21/04/2023
Veppur-606304.	Date of Calibration	: 21/04/2023
,	Due Date of Calibration	: 20/04/2024

Date of Issue

Accuracy

Details of Unit Under Calibration

: Refrigerator

Description

Resolution

	· ·	•	
Make	: LG	Asset ID.No	: NA
Model/ Type	: GL-B191KRLQ	Condition of UUC	: Good
Serial No.	: SADC-004	Calibrated At	: Site
Range	: 2 to 8 °C	Location	: NA

Details of Reference Standard

SI.No	Description	ID.No/Sr.No	Certificate Number	Validity
1	Four wire RTD Sensor	NCS/Thermal/RTD02	TSC/22-23/11472-1	05/11/2023
2	Digital Thermometer	NCS/Thermal/DTM02	TSC/22-23/7495-1	10/08/2023

	Environmental	Temperature (°C)	25 ± 5	Reference Standard	DKD-R-5-7
-	Condition	Humidity (%RH)	55 ± 20	Calibration Procedure	NCS-L03-WI-THER-005
	Calibrated By	A. Naveenkumar			

Calibration Results

Results of Calibration in °C				
SI.No.	Standard Reading (Mean)	UUC Reading (Mean)	Error	Measurement Uncertainty (±)
1	5.250	5	-0.250	0.63

Note:

- 1. The Standards used are traceable to National / International Standards.
- 2. This report refers only to the item calibrated and shall not be reproduced in full without the written approval of the issuing laboratory.
- 3. This report is not valid unless it is signed by the authorized signatory.
- 4. Next Calibration due date mentioned as per customer requirement.
- 5. The result is valid at the time of and under the stated conditions of measurement.
- 6. The measurement confidence level is 95.45% with coverage factor K=2.
- 7. Temperature Scale: ITS-90
- 8. UUC stands for Unit Under Calibration.



Authorised Signatory

Technical Manager

Document Number: NCS-L04-R42 Issue No. & Date: 01 & 19.08.2020