


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230609A/N-115	Page : 1 of 1
CALIBRATED ON. : 10 / 06 / 2023	Ref. Std. : IS:12508:1988
CALIBRATION DUE. : 09 / 06 / 2024	C.P. No. : ACS/CP/10
SRF No. & DATE : 230609A/N Dt. 09.06.2023	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342723000000386F Dt. 14.06.2023	Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009, Bhopal, Madhya Pradesh-462016

ITEM CALIBRATED : Centrifuge
Make : Remi
Range : (0 to 6000) RPM
L. Count :
ID No. : RCL/BIO/Cent-01
Location : Clinical Biochemistry
S. No. : ZIDN-15543

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	11/09/2023

RESULTS : Mechanical Discipline			
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.32	- 1.32	
2000	2002.24	- 2.24	
4000	4005.24	- 5.24	
6000	6006.30	- 6.30	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.50 rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

