

# ADVANCE CALIBRATION SERVICES


(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



CC-3427



## Calibration Certificate

CERTIFICATE NO. : 230609A/N-117 Page : 1 of 1

CALIBRATED ON. : 10 / 06 / 2023 Ref. Std. : CG-11  
CALIBRATION DUE. : 09 / 06 / 2024 C.P. No. : ACS/CP/07

SRF No. & DATE : 230609A/N Dt. 09.06.2023 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342723000000388F Dt. 14.06.2023 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.  
Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009,  
Bhopal, Madhya Pradesh-462016

ITEM CALIBRATED : Digital Temperature Indicator  
Make : RoHS Range : (-50 to 110) °C  
L. Count : ID No. : RCL/BIO/D.Temp-01  
Location : Clinical Biochemistry S. No. : N.A

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C} R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023

RESULTS : ET Discipline		
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)
2.08	2	0.08
4.14	4	0.14
6.16	6	0.16
8.16	8	0.16
(average of readings)		

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

...End of Calibration Certificate...

Auth. Signatory  
S.K. Sharma  
CEO/TM  
C-23, Sec-10, Noida