


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230609A/N-120	Page	: 1 of 1
CALIBRATED ON.	: 10 / 06 / 2023	Ref. Std.	: CG-11
CALIBRATION DUE.	: 09 / 06 / 2024	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 230609A/N Dt. 09.06.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000391F Dt. 14.06.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009, Bhopal, Madhya Pradesh-462016		
ITEM CALIBRATED	: Digital Temperature Indicator		
	Make : RoHS	Range	: (-50 to 110) °C
	L. Count :	ID No.	: RCL/HEM/D.Temp-04
	Location : Haematology	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023

RESULTS : ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.06	2	0.06	
4.08	4	0.08	
6.12	6	0.12	
8.16	8	0.16	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

