


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230609A/N-111	Page	: 1 of 1
CALIBRATED ON.	: 09 / 06 / 2023	Ref. Std.	: DKD R-5-7
CALIBRATION DUE.	: 08 / 06 / 2024	C.P. No.	: ACS/CP/11
SRF No. & DATE	: 230609A/N Dt. 09.06.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000382F Dt. 14.06.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009, Bhopal, Madhya Pradesh-462016		
ITEM CALIBRATED	: Digital Thermo Hygrometer		
	Make : Digital Series	Range	: (0 to 50)°C, (10-99)%
	L. Count :	ID No.	: RCL/CP/HGM-04
	Location : Clical Pathology	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023

RESULTS : Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING		ERROR
(°C)	15.62	14.94	0.68
	25.28	24.96	0.32
	40.62	39.98	0.64
(average of readings)			
(%)	30.58	29.91	0.67
	60.78	59.96	0.82
	90.42	89.94	0.48
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.1°C, 2.5% at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

