


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230609A/N-103	Page	: 1 of 1
CALIBRATED ON.	: 09 / 06 / 2023	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 08 / 06 / 2024	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 230609A/N Dt. 09.06.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000374F Dt. 14.06.2023	Cal. At	: Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009,
Bhopal, Madhya Pradesh-462016

ITEM CALIBRATED	: Micro Pipette	Make	: Thermo Scientific	Range	: (100 to 1000) μ l
		L. Count		ID No.	: RCL/HEM/PIP-03
		Location	: Haematology	S. No.	: TW05923

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	04/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (μ l)	Graduation (μ l)	Nominal Value (μ l)	Volume at 20 °C (μ l)	Error (μ l)
		100	100.50	(-) 0.50
1000		500	501.65	(-) 1.65
		1000	1003.27	(-) 3.27

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.4 \mu$ l at coverage factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

