


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230609A/N-104 Page : 1 of 1

CALIBRATED ON. : 09 / 06 / 2023 Ref. Std. : ISO:8655:6
CALIBRATION DUE. : 08 / 06 / 2024 C.P. No. : ACS/CP/08

SRF No. & DATE : 230609A/N Dt. 09.06.2023 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342723000000375F Dt. 14.06.2023 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
Ground Floor, E-7/137, HIG, Arera Colony, Ward-0049, Zone-0009,
Bhopal, Madhya Pradesh-462016

ITEM CALIBRATED : Micro Pipette
Make : Thermo Scientific Range : (2 to 20) µl
L. Count : ID No. : RCL/BIO/PIP-04
Location : Clinical Biochemistry S. No. : TW08250

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C} R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	04/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.10	(-) 0.10
20		20	20.11	(-) 0.11

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.4 \mu\text{l}$ at coverage factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida