



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 5215

Distribution No.: 158-M

Month/Year: January/2023

Instrument ID: BC-3000plus (RJ8C127453)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 28-02-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	7	6.5	13.5	13.31	0.035	0.22	0.5	0.1	0.007	4.15
RBC x10 <sup>6</sup> /µl	1	5.17	4.97	10.14	10.25	0.012	-0.30	0.2	0.05	0.003	2.89
Hb g/dl	1	12.6	12.2	24.8	26.5	0.027	-2.29	0.4	0.1	0.008	2.02
HCT%	1	37.8	36.5	74.3	85.7	0.219	-1.53	1.3	0.4	0.025	2.31
MCV-fl	1	73.5	73.2	146.7	169.35	0.353	-1.97	0.3	0.3	0.019	0.00
MCH-Pg	1	24.5	24.3	48.8	51.5	0.059	-1.66	0.2	0.2	0.014	0.00
MCHC-g/dl	1	33.4	33.3	66.7	61	0.140	1.05	0.1	0.3	0.021	-0.67
Plt. x10 <sup>3</sup> /µl	1	355	326	681	781	2.986	-1.11	29	9	0.518	2.45
Retic %	2	2.5	2	4.5	15.35	0.222	-1.82	0.5	0.5	0.034	0.00

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbc=0 , Poly=1 L=1, E=1, Mono/Promono=0 , B1=90 P.M.=2, Mye=2, Meta=4, Other=WBC: Total count is increased with shift to left upto myeloblasts.BLASTS MORPHOLOGY: Large cells with moderate cytoplasm,having round to indented nucleus ,coarse chromatin with prominend nucleoli,Auer rods noted in the cytoplasm.PLATELETS: Decreased on

Test Parameters	S.No.	Among Lab (Accuracy Testing)				Within Lab (Precision Testing)				
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values
<b>RBC Morphology</b>	3	Normocytic Normochromic RBC's predominantly with occasional microcytic hypochromic RBC's .Nucleated RBC's noted				Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic				
<b>Diagnosis</b>	3	Normocytic normochromic anemia.Acute leukemia ( acute myeloid leukemia) thrombocytopenia				Acute Leukemia (AL)				

### COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 158--M	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	338	335	83.88	84.18	6.27	6.27	9.85	9.55
<b>RBC x10<sup>6</sup>/µl</b>	1	338	338	89.64	89.35	5.92	5.03	4.44	5.62
<b>Hb g/dl</b>	1	338	338	88.76	85.21	5.03	4.44	6.21	10.35
<b>HCT%</b>	1	338	336	97.92	89.88	0.89	5.36	1.19	4.76
<b>MCV-fl</b>	1	338	336	97.62	87.2	1.79	6.55	0.59	6.25
<b>MCH-Pg</b>	1	338	336	88.39	89.58	7.44	4.76	4.17	5.66
<b>MCHC-g/dl</b>	1	338	336	98.21	86.9	0.89	7.74	0.9	5.36
<b>Plt. x10<sup>3</sup>/µl</b>	1	338	336	94.64	92.26	3.27	2.98	2.09	4.76
<b>ReticCount%</b>	2	338	296	91.22	85.14	6.08	9.8	2.7	5.06
<b>PS Assessment</b>	3	338	283	Satisfactory :97.93%, Borderline Sat. :1.18%, Unsatisfactory :0.890%					

#### Comments:

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Difference in the CBC measurement values for WBC unacceptable, may be due to random/human error.**

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,




Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----





**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 5215

Distribution No.: 157-M

Month/Year: October/2022

Instrument ID: BC-3000plus (RJ8C127453)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 24-11-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.4	5.2	10.6	11.2	0.029	-0.74	0.2	0.1	0.006	0.96
RBC x10 <sup>6</sup> /µl	1	3.73	3.73	7.46	7.55	0.008	-0.42	0	0.04	0.003	-0.90
Hb g/dl	1	10.9	10.8	21.7	23.7	0.027	-3.03	0.1	0.1	0.008	0.00
HCT%	1	31	30.9	61.9	73.3	0.166	-2.37	0.1	0.4	0.025	-0.58
MCV-fl	1	83.3	82.9	166.2	194.6	0.396	-2.46	0.4	0.3	0.021	0.34
MCH-Pg	1	29.2	28.9	58.1	62.6	0.084	-1.90	0.3	0.3	0.020	0.00
MCHC-g/dl	1	35.1	34.9	70	64.5	0.150	1.30	0.2	0.3	0.022	-0.27
Plt. x10 <sup>3</sup> /µl	1	143	129	272	281	1.191	-0.28	14	4	0.275	2.25
Retic %	2	1.7	1	2.7	10.5	0.232	-1.18	0.7	0.5	0.033	0.34

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3 Nrbcs=2 , Poly=60 L=1, E=8, Mono/Promono=1 , B1=8 P.M.=2, Mye=10, Meta=8, Other=PLATELETS ADEQUATE ON SMEAR	Poly: 60 - 77, Myelo: 5 - 12, Meta: 5 - 10, Lympho: 3 - 7, Eos: 1- 3, nRBC/ Baso/ Promyelo, Blast Mono: 0 - 5
RBC Morphology	3 NORMOCYTIC IN NORMOCHROMIC ,RBCs WITH OCCASIONAL TEAR DROP CELLS AND FEW PENCIL CELLS .NRBCs SEEN	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia
Diagnosis	3 CHRONIC MYELOID LEUKEMIA	Chronic Myeloid Leukemia

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 157--M	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	334	333	83.18	88.59	6.61	5.11	10.21	6.3
<b>RBC x10<sup>6</sup>/µl</b>	1	334	334	88.62	88.92	5.09	5.69	6.29	5.39
<b>Hb g/dl</b>	1	334	334	86.53	85.93	5.99	6.89	7.48	7.18
<b>HCT%</b>	1	334	332	93.98	91.57	4.22	3.31	1.8	5.12
<b>MCV-fl</b>	1	334	333	95.5	90.99	3	2.4	1.5	6.61
<b>MCH-Pg</b>	1	334	333	90.09	85.59	5.71	7.81	4.2	6.6
<b>MCHC-g/dl</b>	1	334	333	93.69	91.89	3.9	2.1	2.41	6.01
<b>Plt. x10<sup>3</sup>/µl</b>	1	334	333	91.29	91.89	5.71	4.2	3	3.91
<b>ReticCount%</b>	2	334	297	87.88	88.22	7.41	7.07	4.71	4.71
<b>PS Assessment</b>	3	334	270	Satisfactory :87.66%, Borderline Sat. :11.14%, Unsatisfactory :1.20%					

**\*Comments:**

- 1). Among Lab (EQA) : CBC result for HB unacceptable, may be due to random/human error**
- 2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----