



CC - 3237



#40, 2nd Floor, 2nd Street,
Padmavathi Nagar, Chromepet,
Chennai, Tamil Nadu - 600 044.

CALIBRATION CERTIFICATE

MCAL/FOR/QD-12

Page # 1 of 1

| | | | |
|---------------|-----------------------|---------------------|---------------|
| PRF # | : MCAL/PRF/0623189 | Date of Calibration | : 16-Jun-2023 |
| Certificate # | : MSIR/410623/189-003 | Recom. Due Date | : 15-Jun-2024 |

DETAILS OF USER CUSTOMER

M/S. SHIFA HOSPITAL,

#59, Pattukottai Road, Adirampattinam,

Tanjavur Dist. Tamil Nadu-614 701

| | |
|------------------|----------------------|
| ULR # | : CC323723000008443F |
| Calibrated at | : On-site |
| Date of Receipt | : 16-Jun-2023 |
| Cond. On Receipt | : Satisfactory |
| Date of Issue | : 22-Jun-2023 |

DETAILS OF TEST / UNIT UNDER CALIBRATION

| | | | |
|----------------------|---------------------|------------------|------------------|
| Description | : Micropipette | Identification # | : MP-01 |
| Range | : 10 to 100 μ l | Serial # | : YE189AL0094972 |
| Cal. / Working Range | : -- | Make | : GRAGON LAB |
| Least Count | : 1 μ l | Model # / Input | : -- |
| Location | : -- | Accuracy | : -- |

DETAILS OF MASTER INSTRUMENT USED

| Description (Range) | Identification # | Valid upto (Due Date) | Traceability Certificate # |
|---------------------|------------------|-----------------------|----------------------------|
| Weighing Balance | MCAL/INS/MV-04 | 16-Dec-2023 | MSIR/411222/168-001 |

| | |
|-----------------|---|
| MFC # | : MCAL/MFC/MV-03 |
| Details of Env. | : Temperature : 25 \pm 2 $^{\circ}$ C Relative Humidity : 50 \pm 10 % RH |

Remarks

1. The reported Expanded Uncertainty is calculated at 95.45 % C.L with coverage factor k=2.
2. DUC is defined as Device Under Calibration.

Issued by :
MCAL - Engineer

SANGEETHA. B

Authorised by :
Technical Manager

KAMALANATHAN. V

MSIR INDIA PRIVATE LIMITED

(The Madras Science and Industrial Resources)

Phone :044 2265 2634, +91 9789877567, Email: calibration@msirindia.com Web: www.msirindia.com