



Accumen Calibration Laboratory Pvt Ltd.

(Calibration Division)

B-20, Hanumant Vihar, Naubasta, Kanpur Nagar-208021 (U.P)

Mob.: +91-8802463273, 8417800601, 8750869237

E-mail:- accumencalibration22@gmail.com



CC-3513

CALIBRATION CERTIFICATE

Format No.: 7.8F-01

Page 1 of 1

ULR No. : CC351323000000475F	Certificate No. : ACL/004375
CUSTOMER NAME & ADDRESS:- M/s. Photon Super Speciality Hospital and Trauma centre. Kashipur Road Rampur Doraha Opp., Radha Swami Satsang Bhawan Moradabad (U.P)-244001.	Service Request No : 0321
	Service Request Date. : 28-02-2023
	Date of Receipt of UUC : 01-03-2023
	Date of Calibration : 01-03-2023
	Sugg. Due Date By Customer : 28-02-2024
	Date of Issue : 02-03-2023

Instrument Name : CNTRIFUGE	Range/Size : Ref. to Result.
Make : Polestar	Least Count : ----
Model No. : ---	Visual Inspection : OK
Serial No. : ABMT1605681122	Location : ----
Party ID Mark : ---	

ENVIRONMENTAL CONDITIONS	Calibration Procedure : MCP-23
Temperature : (25 ± 10)°C	Reference Standard : IS 12508 : 2013
Relative Humidity : (50 ± 15)%	Calibration Performed at : Site

MAJOR EQUIPMENT USED				
Nomenclature	Range	Serial/Ident. No.	Traceability Certificate No.	Valid upto
Digital tachometer (Contact /Non Contact)	up to 99999.9RPM	204-0014V/ACL/MMI/07	QCTS/000077/04	05-04-2023

CALIBRATION RESULT

DUC Set Step	Observed Value in Standard (in RPM)	Uncertainty of Measurement (±)%
1	499.3	0.12
2	1000.9	0.12
3	1501.2	0.12
4	2002.0	0.12
5	2503.1	0.12

Remarks:

Note :

- DUC = Device Under Calibration, UUC=Unit Under Calibration
- Expanded uncertainty of measurement at approx 95% confidence level for coverage factor(k)=2
- This Certificate refers only to the particular items submitted for calibration.
- Suggested next due date of calibration is given as requested by customer.
- DUC has been calibrated against lab standards are traceable to national & international standards.
- The result reported in this certificate are valid at the time of and under the stated conditions of Measurement.
- This certificate shall not be reproduced, except in full, without permission of CEO, Accumen Calibration Laboratory Pvt. Ltd., Kanpur (U.P.)

Issue by

Manish Verma
(Calibrated By)



End of Report

Jaysagar yadav
(TM/QM)