



Accumen Calibration Laboratory Pvt Ltd.

(Calibration Division)

B-20, Hanumant Vihar, Naubasta, Kanpur Nagar-208021 (U.P)

Mob.:+91-8802463273, 8417800601, 8750869237

E-mail:- accumencalibration22@gmail.com



CALIBRATION CERTIFICATE

Format No.:7.8F-01

Page 1 of 1

ULR No. : CC351323000000474F	Certificate No. : ACL/004374
CUSTOMER NAME & ADDRESS:- M/s. Photon Super Speciality Hospital and Trauma centre. Kashipur Road Rampur Doraha Opp., Radha Swami Satsang Bhawan Moradabad (U.P)-244001.	Service Request No : 0321
	Service Request Date. : 28-02-2023
	Date of Receipt of UUC : 01-03-2023
	Date of Callibration : 01-03-2023
	Sugg. Due Date By Customer : 28-02-2024
	Date of Issue : 02-03-2023

ITEM'S DESCRIPTION & IDENTIFICATION

Instrument Name : KITS REFRIGERATED (DTI WITH INDICATOR)	Range/Size : 2 to 8 °C
Make : Godrej	Least Count : 0.1 °C
Model No. : R01EDG	Visual Inspection : OK
Serial No. : -----	Location : -----
Party ID Mark : RF-01	

ENVIRONMENTAL CONDITIONS	Calibration Procedure : TCP-03
Temperature : (25 ± 10)°C	Reference Standard : DKD-R-5-1
Relative Humidity : (60 ± 15)%	Calibration Performed at : Site

MAJOR EQUIPMENT USED				
Nomenclature	Range	Serial/Ident. No.	Traceability Certificate No.	Valid upto
Digital Thermometer with 4 wire RTD Sensor	-80 to 300 °C	21J668964/22000027 ACL/MTI/02	TSC/21-22/16764-1	22-02-2023

CALIBRATION RESULT


Measured Value in DUC* (in °C)	Observed Value in Std. (in °C)	Error ± (in °C)	Uncertainty of Measurement (±)
2.0	1.896	0.104	0.38°C
4.0	3.887	0.113	0.38°C
6.0	5.878	0.122	0.38°C

Remarks:

- Note :
- DUC = Device Under Calibration, UUC=Unit Under Calibration
 - Expanded uncertainty of measurement at approx 95% confidence level for coverage factor(k)=2
 - This Certificate refers only to the particular items submitted for calibration.
 - Suggested next due date of calibration is given as requested by customer.
 - DUC has been calibrated against lab standards are traceable to national & international standards.
 - The result reported in this certificate are valid at the time of and under the stated conditions of Measurement.
 - This certificate shall not be reproduced, except in full, without permission of CEO, Accumen Calibration Laboratory Pvt. Ltd., Kanpur (U.P.)

Issue by

Satendra Singh
(Calibrated By)



End of Report

Jaysagar yadav
(TM/QM)