









C-31, VMB Tower, Sector-13, Vasundhra, Ghaziabad, U.P. 201012

Good

Mobile: +91-9999788038, Tel.: 0120-4919433 E-Mail: technomeasurepvt.ltd@gmail.com

CERTIFICATE OF CALIBRATION

Date of Receipt Item:

Cert. Issue Date:

Date of Calibration:

01.03.2023

ULR No.:

CC282123000005605F

03.03.2023

Format No./REV. No./REV. DATE: 7.8F-01/03/01.06.2022

01.03.2023

Certificate No.:

Condition of the Item:

TMPL/23/05605

Next Date for Calibration:

29.02.2024

Page No. 1 of 1

(As per agreed with the Customer)

		(in particular transfer and tra		
CUSTOMER ADDRESS		EQUIPMENT DESCRIPTION & IDENTIFICATION		
M/s. PHOTON SUPER SPEC!ALITY HOSPITAL AND TRAUMA CENTRE		Nomenclature :	Microscope Binocular	
KASHIPUR ROAD RAMPUR DORAHA OPP.RADHA SWAMI SATSANG BHAWAN MORADABAD (U.P)-244001		Range :	5X, 10X, 40X & 100X.	
		Least Count :	N/A	
<u></u>		Make /Model No:	LABOMED/LABOMED-2000	
CALIBRATION REQUEST DETAILS		Serial No. :		
Service Request No.:	TMPL/23/0334	I/D Mark:		
Dated :	01.03.2023	Location:		

MASTER EQUIPMENT USED

S. No	Inst. Name	Cert. No.	Cal. By	Calibration Date	Cal. Due Date
1	Glass Scale	TI/C/GS/006/21	TANSON	27.03.2021	26.03.2023
2	Glass Scale	TI/F/GS/024/22	TANSON	24.06.2022	23.06.2023

ENVIRONMENTAL CONDITIONS (AT SITE)		PROCEDURE FOLLOWED	REFERENCE STANDARD
Temperature	: 25±10°C	CP/T/S-01	111111111111111111111111111111111111111
Humidity	: 55±15%		
Discipline Mech	anical		

CALIBRATION RESULT

Eye Piece x Objective = Magnification	Reference Size in (mm)	Observed Magnification in (mm)
10x X 5x = 50x	0.5	25.00
10x X 10x =100x	0.5	50.00
10x X 40x = 400x	0.2	79.99
10x X 100x = 1000x	0.1	99.99

Note: *UUC: Unit Under Calibration

Uncertainty of measurment at approx 95% Confidence Level with coverage factor k = 2, is \pm 0.4 %

*** End of Certificat

Calibrated BY

Raj Kumar Prajapati Technical Manager **CHECKED BY**

Dhananjay Mishra **Quality Manager** APPROVED BY

1. The Certificate refers only to the particular item submitted for Calibration.

2. The Calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement.

3. This particular certificate shall not be reproduced except in full, without prior permission of Chief Executive Officer of the Lab.