



April 13, 2023

Calibration Certificate for Semi Automated Biochemistry Analyser, Model: ERBA Chem 7 (SLNo. – S1901148)

This is to certify that Semi Automated Biochemistry Analyser, Model: Erba Chem -7 installed at MediQ Diagnostic Center Ho-chi-Minh Sarani (1st Floor) Sukuntala Park, Sarsuna, Kolkata 700 061, West Bengal, has been calibrated and the calibration value of filter gains and all other values has been checked and found well within the limit.

This calibration is valid till April 12, 2024

For Transasia Bio-Medicals Ltd.

*Suman Senapati
Service Engineer*



TECHNICAL SERVICE REPORT

No. **1493293**

DATE: **13/04/23**

CUSTOMER DETAILS	INSTRUMENT DETAILS	SERVICE STATUS
NAME: Medip Diagnostic centre.	MODEL: chem 7	<input type="checkbox"/> WARRANTY <input type="checkbox"/> R&R <input type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> CHARGED CALL
ADDRESS: Sarewan, Kolkata	SR. NO.: 31901148	
CALL DETAILS		TYPE OF CALL
	DATE: TIME:	<input type="checkbox"/> INSTALLATION
COMPLAINT RECD.		<input type="checkbox"/> P.M. VISIT <input type="checkbox"/> I <input type="checkbox"/> II <input type="checkbox"/> III <input type="checkbox"/> IV
RESPONSE	13/4	<input type="checkbox"/> APPLICATION SUPPORT
JOB COMPLETED	13/4	<input type="checkbox"/> BREAKDOWN
RESPONSE TIME		DOWN TIME :
TRAVEL TIME		COUNTER READING :

PROBLEM REPORTED: **Calibration**

OBSERVATIONS: **-DO-**

ACTION TAKEN: **checked operating environment (power supply, temperature, humidity). found operating environment is satisfactory. checked the adapter o/p voltage and all hardware function. all are ok. Necessary calibration done. working well.**

SITE CONDITION: LINE-NEUTRAL VOLT.: **230V** NEUTRAL-EARTH VOLT.: **1V** LINE-EARTH VOLT.: **229V**

BRAND OF REAGENT USED :					TO BE FILLED IN BY CUSTOMER	
<input type="checkbox"/> FOLLOWING PARTS HAVE BEEN REPLACED <input type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED. PLEASE APPROVE					<input type="checkbox"/> PREVENTIVE MAINTENANCE CARRIED OUT SATISFACTORILY. <input type="checkbox"/> FAULT RECTIFIED & INSTRUMENT IS WORKING SATISFACTORILY. <input type="checkbox"/> WE HEREBY APPROVE RS. _____ FOR PARTS <input type="checkbox"/> COMMENTS (IF ANY) :	
NO.	DESCRIPTION	QTY.	COST	TOTAL		
/					SEAL DATE CUSTOMER'S SIGNATURE NAME :	
TOTAL Rs.						

INVOICE NO. :	DATE :	BRANCH	H. O.
FOLLOW-UP ACTION (Required if any) :		RECEIVED ON :	CHECKED BY :

ENGINEER'S/APPLICATION SPECIALIST'S SIGNATURE : **Suman Senapat**
 TIME : NAME :

NOTE: Parts replaced are chargeable except during warranty period. Consumables like printer head, lamp, tubing, paper rolls etc. & breakable parts are not covered by warranty and hence are chargeable. Parts replaced due to negligence in operation will also be charged in every case.
AT TRANSASIA, CUSTOMER SATISFACTION IS OUR PRIME CONCERN. IN CASE YOU HAVE ANY SUGGESTIONS PLEASE CONTACT : GENERAL MANAGER (TECHNICAL SERVICE), MUMBAI. TEL. : 4030 9000

TRANSASIA BIO-MEDICALS LTD.
 Head Office : Transasia House, 8, Chandivali Studio Road, Andheri (E), Mumbai - 400 072. Tel. : (022) 4030 9000 Fax : (022) 2857 3030
 Mumbai : Toll Free No. 1800 103 8226, Whatsapp No. : 8451048434, SMS No. : 9212433444
 Indore : Tel. : (0731) 403 3948 Chennai : Tel. : (044) 2822 7149 Kolkata : Tel. : (033) 2215 7839 Fax : 2215 8035
 Lucknow : Tel. : (0522) 443 0900 Pune : Tel. : (020) 2616 2658 Dehi : Tel. : (011) 2578 5451/2573 2223
 Chandigarh : Tel. : (0172) 464 3837 Hyderabad : Tel. : (040) 4430 9000 Ahmedabad : Tel./Fax : (079) 02640 7030/4032 1903
 Jaipur : Tel. : 93092 63147 Guwahati : Tel. : 91369 50480 Bangalore : Tel./Fax : (080) 2556 8044
 Patna : Tel. : 77669 00553 Chuttack : Tel. : (0671) 232 4555
 Kochi : Tel. : (0484) 402 0511 Hubli : Tel. : (0836) 485 0900