

# Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307  
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com  
 Website : www.pmmpl.in, www.parashar.info



The way to Technology  
 (20 Years of Excellence)



CC - 2795


## Calibration Certificate

<b>ULR - CC279523000032981F</b>			<b>Page 1 of 1</b>		
<b>Certificate Number : PMM/140623/16-1</b>			<b>Date of Request</b>		
<b>Calibrated For</b>		M/s. Department Of Pathology Tej Bahadur Sapru Hospital, Prayagraj.		<b>Calibrated on</b>	
				14.06.2023	
				<b>Valid Upto</b>	
				13.06.2024	
				<b>Cert. Issue Date</b>	
				16.06.2023	
<b>Description of Equipments</b>					
<b>Nomenclature</b>		Centrifuge		<b>Accuracy/Class</b>	
Range		500 to 3000 rpm		Not Mentioned	
<b>Resolution / Least Count</b>		10 rpm		<b>Condition of UUC</b>	
Make & Model No.		Laby / T3 DLD		Physically Ok	
<b>Material / Type</b>		-----		<b>Serial No.</b>	
				2017-18/133	
				<b>Party ID Mark No.</b>	
				TBSH/DP/CF-01	
				<b>Location</b>	
				Pathology	
<b>Master Equipment / Standard Used</b>					
<b>Sr.No.</b>	<b>Nomenclature</b>	<b>Make &amp; Model</b>	<b>Serial No. / I/D Mark</b>	<b>Calibrated From</b>	<b>Certificate No.</b>
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMP, Noida.	PMM/1418
<b>Due Date of Calibration</b>					
04.12.2023					
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration					
<b>Calibration Performed at</b>		<b>Site</b>		<b>Environmental</b>	
Calibration Procedure No.		CP/M/MV/S/03		Condition(s)	
Reference Standard / Guideline		-----		Temperature = 23 ± 5° C	
				Humidity = 30 to 75 % RH	
<b>Mechanical Calibration</b>					
<b>CALIBRATION RESULTS</b>					
<b>Sr.No.</b>	<b>Nominal Value in UUC (in rpm)</b>	<b>Actual Observed Value in Std. (in rpm)</b>		<b>Error (in rpm)</b>	<b>± Uncertainty in Measurement (in rpm)</b>
1	500	501		-1	51
2	1000	1002		-2	51
3	1500	1503		-3	51
4	2000	2004		-4	51
5	3000	3005		-5	51


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor  $k = 2$ )

Calibrated By :  Sudhir Kumar  
 Designation : Calibration Engineer



Approved By :  Jai Prakash  
 Designation : Tech. Manager / Auth. Sign.

- Conditions :**
1. This certificate refers only to the particular item submitted for calibration.
  2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
  3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.