



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230612A/N-115	Page	: 1 of 1	
CALIBRATED ON.	: 13 / 06 / 2023	Ref. Std.	: IS:12508:1988-	
CALIBRATION DUE.	: 12 / 06 / 2024	C.P. No.	: ACS/CP/10	
SRF No. & DATE	: 230612A/N Dt. 12.06.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342723000000407F Dt. 14.06.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 03, Civic Center, Main New Road, Ratlam-457001			
ITEM CALIBRATED	: Centrifuge			
	Make : Remi	Range	: (0 to 6000) RPM	
	L. Count :	ID No.	: RCL/BIO/Cent-01	
	Location : Clinical Biochemistry	S. No.	: ZIEN-24031	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	11/09/2023
RESULTS	Mechanical Discipline			
	* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
	1000	1001.30	- 1.30	
	2000	2002.26	- 2.26	
	4000	4005.34	- 5.34	
	6000	6006.20	- 6.20	
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.50 rpm of reading at coverage Factor $k=1.96$.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

