


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 230612A/N-119	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 13 / 06 / 2023	<b>Ref. Std.</b> : CG-11
<b>CALIBRATION DUE.</b> : 12 / 06 / 2024	<b>C.P. No.</b> : ACS/CP/07
<b>SRF No. &amp; DATE</b> : 230612A/N Dt. 12.06.2023	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342723000000411F Dt. 14.06.2023	<b>Cal. At</b> : Site / Lab

<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. 03, Civic Center, Main New Road, Ratlam-457001
--

<b>ITEM CALIBRATED</b> : Digital Temperature Indicator	
<b>Make</b> : RoHS	<b>Range</b> : (-50 to 110) °C
<b>L. Count</b> :	<b>ID No.</b> : RCL/BIO/D.Temp-03
<b>Location</b> : Clinical Biochemistry	<b>S. No.</b> : N.A

<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. {40% to 60%}
--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023

RESULTS : ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.08	2	0.08	
4.08	4	0.08	
6.14	6	0.14	
8.16	8	0.16	
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory  
S.K. Sharma  
CEO/TM  
C-23, Sec-10, Noida

....End of Calibration Certificate....