


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 230612A/N-107	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 12 / 06 / 2023	<b>Ref. Std.</b>	: DKD R-5-7.	
<b>CALIBRATION DUE.</b>	: 11 / 06 / 2024	<b>C.P. No.</b>	: ACS/CP/11	
<b>SRF No. &amp; DATE</b>	: 230612A/N Dt. 12.06.2023	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342723000000399F Dt. 14.06.2023	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. 03, Civic Center, Main New Road, Ratlam-457001			
<b>ITEM CALIBRATED</b>	: Digital Thermo Hygrometer			
	<b>Make</b> : Digital Series	<b>Range</b>	: (0 to 50)°C, (10-99)%	
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/HEM/HGM-01	
	<b>Location</b> : Haematology	<b>S. No.</b>	: N.A	
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023
<b>RESULTS</b>	: Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR		
(°C)	15.22	14.96	0.26	
	25.26	24.97	0.29	
	40.24	39.95	0.29	
(average of readings)				
(%)	30.06	29.95	0.11	
	60.26	59.96	0.30	
	90.24	89.94	0.30	
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.1°C, 2.5% at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

