


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230612A/N-110	Page : 1 of 1			
CALIBRATED ON. : 12 / 06 / 2023	Ref. Std. : DKD R-5-7			
CALIBRATION DUE. : 11 / 06 / 2024	C.P. No. : ACS/CP/11			
SRF No. & DATE : 230612A/N Dt. 12.06.2023	Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342723000000402F Dt. 14.06.2023	Cal. At : Site / Lab			
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. 03, Civic Center, Main New Road, Ratlam-457001				
ITEM CALIBRATED : Digital Thermo Hygrometer				
Make : Digital Series	Range : (0 to 50)°C, (10-99)%			
L. Count :	ID No. : RCL/SC/HGM-04			
Location : Sample Collection Room	S. No. : N.A			
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023
RESULTS : Thermal Discipline				
* U. C. INSTT. READS	STANDARD READING		ERROR	
(°C)	15.58	14.92	0.66	
	25.50	24.96	0.54	
	40.50	39.94	0.56	
(average of readings)				
(%)	30.42	29.94	0.48	
	60.46	59.95	0.51	
	90.60	89.93	0.67	
(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.1°C, 2.5% at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

