


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230612A/N-102	Page	: 1 of 1
CALIBRATED ON.	: 12 / 06 / 2023	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 11 / 06 / 2024	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 230612A/N Dt. 12.06.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000394F Dt. 14.06.2023	Cal. At	: Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
03, Civic Center, Main New Road, Ratlam-457001

ITEM CALIBRATED	: Micro Pipette		
Make	: Thermo Scientific	Range	: (10 to 100) µl
L. Count	:	ID No.	: RCL/HEM/PIP-02
Location	: Haematology	S. No.	: TW08036

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C} R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	04/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.17	(-) 0.17
100		50	50.21	(-) 0.21
		100	100.38	(-) 0.38

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmitly statement not provided since not required.

Prepared by *Bh*

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida