

# INSTALLATION REPORT

# BioSystems

**CUSTOMER** INSTALLATION DATE: 30/04/2023  
Customer Name: HOD BIOCHEMISTRY, BHU/Prof. Incharge CCI Lab.  
Address: CCD LAB, SIRSUNDELLAL HOSPITAL  
BHU, VARANASI  
UP-221005  
Contact Person: Dr. S.P. MISHRA  
Phone: Mobile: 9453172674  
E-mail:

**INSTRUMENT DETAILS / STATUS**  
MODEL NAME: BA400 SERIAL NUMBER: 834002703  
SOFTWARE VERSION: 5.9.1.0  
1) Was there any external damage on the instrument or accessories? Yes/No ✓  
2) Were any of the accessories missing? Yes/No ✓  
3) Were any parts loose? Any unsecured hardware? Yes/No ✓  
If any of the above were answered YES, please elaborate

**PRE-INSTALLATION CHECKS**  
1) is the room air-conditioned? Yes/No ✓  
2) Vibration free platform? Yes/No ✓  
3) Mains A.C. Supply  CVT  U.P.S  STABILIZER  
specified Capacity & Make: 20KVA  
Output Voltage: 226V Line & Neutral 226V  
Neutral & Earth Line & Earth: 0-1V, 227V

The above mentioned instrument has been satisfactorily installed by the Engineer / Product Specialist of BioSystems Diagnostics Pvt. Ltd.,  
Received all standard accessories in good condition  
Customer's Remarks:  
Name & Signature of Customer: [Signature] 30/4/23  
Name & Signature of Engineer: ROHIT SINGH  
Date & Place: 17/05/2023 VARANASI  
Seal & Date: S.S. Hospital, I.M.S., B.H.U.

**Representative Office:**  
BioSystems Diagnostics Pvt. Ltd.  
Factory Address:  
A4, SIPCOT Industrial Park, Irungattukottai, Thandalam Village, Sriperumbudur - 602105.  
Tel +91-44-3034 1911 / 12 / 13  
e-mail: sat@biosystems.in

**Corporate Office:**  
BioSystems S.A.  
Costa Brava 30, 08030 Barcelona (Spain)  
Tel +34-93 311 08 11  
Fax +34-93 346 79 99

<b>User Training on Instrument Operation</b>	
Operator Name: Abhishek Tripathi	
Trainer Name: ROHIT SINGH	
Training Date: 17/05/2023	
Operator Sign: [Signature]	Trainer Sign: [Signature]