

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(20 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



CC - 2795

Calibration Certificate

ULR - CC279523000033434F			Page 1 of 1			
Certificate Number : PMM/170623/7-2			Date of Request		17.06.2023	
Calibrated For		M/s.Center For Clinical Investigations Sir Sunderlal Hospital, Institute Of Medical Science, Banaras Hindu University, Varanasi.	Calibrated on		17.06.2023	
			Valid Upto		16.06.2024	
			Cert. Issue Date		19.06.2023	
Description of Equipments						
Nomenclature		Dig. Centrifuge	Accuracy/Class		Not Mentioned	
Range		0 - 4200 rpm	Condition of UUC		Physically Ok	
Resolution / Least Count		5 rpm	Serial No.		SB082672	
Make & Model No.		Neutron / IfugeL 1000	Party ID Mark No.		CF - 02	
Material / Type		-----	Location		Lab	
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMP, Noida.	PMM/1418	04.12.2023
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site	Environmental Condition(s)		Temperature = 23 ± 5° C	
Calibration Procedure No.		CP/MMV/S/03			Humidity = 30 to 75 % RH	
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)		Error (in rpm)	± Uncertainty in Measurement (in rpm)	
1	1000	1001		-1	51	
2	2000	2002		-2	51	
3	3000	3002		-2	51	
4	4000	4003		-3	51	
5	4200	4205		-5	51	


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : 
Designation : Calibration Engineer



Approved By : 
Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.