



Alere Medical Pvt. Ltd.  
No. 404, BPTP Park Centra,  
NH - 8, Sector - 30  
Opp. 32nd Milestone, Gurugram  
Haryana 122001, India  
T: +91 124 4569000  
F: +91 124 4569011

### Calibration Certificate

**Instrument Name** : Urometer120  
**Serial Number** : 12076624  
  
**Customer Name** : Sri Siva Health Care Unit(A Unit Of Tesla Diagnostics)  
**Address** : karmanghat, Hyderabad  
  
**Standard used** : Self Calibration Mode  
**Standard Traceability** : In House Certificate

#### Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of  $333_{\pm 10}$ .

QC Performance : Acceptable  
Self Calibration : Acceptable

**Performance Qualification : Passed**

Calibration Date : 08.02.2023  
**Calibration** : **Passed**  
Calibration Valid up to :07.02.2024

**For Alere Medical Pvt Ltd.**

Upendrakumar V  
Engineer Technical Services  
M: 9052723178

