

## **Calibration Certificate**

Instrument Name

: Urometer120

Serial Number

: 12076626

**Customer Name** 

: Sri Vishnu Health Care Unit(A Unit Of Tesla Diagnostics)

Address

: Sangareddy

Standard used

: Self Calibration Mode

Standard Traceability

: In House Certificate

## Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333±10.

QC Performance

: Acceptable

Self Calibration

: Acceptable

Performance Qualification: Passed

Calibration Date

: 20.04.2023

Calibration

: Passed

Calibration Valid up to

:19.04.2024

For Abbott Diagnostics Medical Pvt Ltd.

Upendrakumar V

VIII

**Engineer Technical Services** 

M: 9052723178