


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 230620B/N-101	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 20 / 06 / 2023	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 19 / 06 / 2024	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DATE</b>	: 230620B/N Dt. 20.06.2023	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342723000000539F Dt. 22.06.2023	<b>Cal. At</b>	: Site / Lab

<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Shop No-203, 2 <sup>nd</sup> Floor Landmark Complex, Prahlad Nagar 100ft Road, Near Seema Hall, Ahmedabad
-----------------------	--

<b>ITEM CALIBRATED</b>	: Micro Pipette
<b>Make</b>	: Thermo Scientific
<b>Range</b>	: (2 to 20) µl
<b>L. Count</b>	: ID No. : RCL/HEM/PIP-01
<b>Location</b>	: Haematology
	: S. No. : SH88098

<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }
-------------------------------	--

<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	05/01/2024

<b>RESULTS :</b>				
Mechanical Discipline				
<b>Nominal Capacity (µl)</b>	<b>Graduation (µl)</b>	<b>Nominal Value (µl)</b>	<b>Volume at 20 °C (µl)</b>	<b>Error (µl)</b>
		10	10.10	(-) 0.10
20		20	20.14	(-) 0.14

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.02 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

