


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230620B/N-102	Page	: 1 of 1	
CALIBRATED ON.	: 20 / 06 / 2023	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 19 / 06 / 2024	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 230620B/N Dt. 20.06.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342723000000540F Dt. 22.06.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Shop No-203, 2 nd Floor Landmark Complex, Prahlaad Nagar 100ft Road, Near Seema Hall, Ahmedabad			
ITEM CALIBRATED	: Micro Pipette			
	Make : Thermo Scientific	Range	: (20 to 200) µl	
	L. Count	ID No.	: RCL/HEM/PIP-02	
	Location : Haematology	S. No.	: SH78463	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	05/01/2024
RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		20	20.15	(-) 0.15
200		100	100.40	(-) 0.40
		200	200.72	(-) 0.72

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.03 µl at coverage factor k=1.96.

- *U.C. : Under Calibration.
- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
 - The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
 - This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
 - The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
 - Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
 - Conformity statement not provided since not required.

Prepared by

...End of Calibration Certificate...

