

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



Calibration Certificate

ULR - CC279523000033534F			Page 1 of 1			
Certificate Number : PMM/170623/10-1			Date of Request			
Calibrated For			Calibrated on			
Department Of Pathology Rani Laxmi Bai Combined Hospital, Rajajipuram, Lucknow.			17.06.2023 17.06.2023 16.06.2024 20.06.2023			
Description of Equipments						
Nomenclature		Dig. Centrifuge		Accuracy/Class		
Range		500 to 4500 rpm		Not Mentioned		
Resolution / Least Count		10 rpm		Condition of UUC		
Make & Model No.		Remi / R - 8C Plus		Physically Ok		
Material / Type		-----		Serial No.		
				ZHKN - 43376		
				Party ID Mark No.		
				CF - 01		
				Location		
				Pathology		
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMPL, Noida.	PMM/1418	04.12.2023
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site		Environmental		Temperature = 23 ± 5° C
Calibration Procedure No.		CP/MMV/S/03		Condition(s)		Humidity = 30 to 75 % RH
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)		Error (in rpm)	± Uncertainty in Measurement (in rpm)	
1	1000	1000		0	51	
2	2000	1999		1	51	
3	3000	2998		2	51	
4	4000	3997		3	51	
5	4500	4496		4	51	

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : Sudhir Kumar
Designation : Calibration Engineer



Approved By : Jai Prakash
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.