

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



The way to Technology
(20 Years of Excellence)



Calibration Certificate

ULR - CC279523000033535F		Page 1 of 1				
Certificate Number : PMM/170623/10-2		Date of Request	17.06.2023			
Calibrated For	Department Of Pathology Rani Laxmi Bai Combined Hospital, Rajajipuram, Lucknow.	Calibrated on	17.06.2023			
		Valid Upto	16.06.2024			
		Cert. Issue Date	20.06.2023			
Description of Equipments						
Nomenclature	Dig. Centrifuge	Accuracy/Class	Not Mentioned			
Range	400 - 2800 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	10 rpm	Serial No.	2017 - 18 / 125			
Make & Model No.	Laby / T3 BLD	Party ID Mark No.	CF - 02			
Material / Type	-----	Location	Pathology			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMPL, Noida.	PMM/1418	04.12.2023
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site	Environmental Condition(s)	Temperature = 23 ± 5° C		
Calibration Procedure No.		CP/M/MV/S/03		Humidity = 30 to 75 % RH		
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)		
1	500	500.8	-0.8	51		
2	1000	1001	-1	51		
3	2000	2001	-1	51		
4	2500	2502	-2	51		
5	2800	2803	-3	51		

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Sudhir

Calibrated By : Sudhir Kumar
Designation : Calibration Engineer



Jai Prakash

Approved By : Jai Prakash
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.