

MEASURE TECHNO LAB

2, B.T. ROAD (JAYANTI CINEMA COMPLEX), BARRACKPORE,
KOLKATA - 700120, W.B.

Ph: 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974,
LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in



CC - 2545

CALIBRATION CERTIFICATE OF MICRO PIPETTE

Certificate No.: MTL / KSDH / R03 / 04 - 23

ULR - CC254523000008968F

Page: 1 of 1

Request No.: MTL / 05 / 04 / 23 - 24

HIV Testing Laboratory,
ICTC Unit,
Kharagpur Sub Division Hospital,
Kharagpur, Paschim Medinipur,
Pin - 721301.

Instrument &
Description of item
Calibrated:

a) Name: Micro Pipette
b) Code No.: KSDH / ICTC / MP - 02
c) Sl. No.: 19227538
d) Make: Microlit
e) Model / Type: N.S.
f) Range: 5 µl to 50 µl
g) Sensor: N.A.
h) Resolution: 1 µl
i) End User: Laboratory
j) Accuracy: N.S.

Date of receipt of item : 04-04-23
Date of calibration : 05-04-23
Date of issue : 07-04-23

k) Calibration done at: On Site / In House
1.4 Physical Condition of DUC : OK
1.6 Recommended date of next calibration : 05-04-24

Environmental Conditions During Calibration:

Temperature: 20 °C ± 2 °C
Humidity: 30 % RH to 75 % RH
Pressure: 1005.3 mbar

Method of Calibration:

SOP / MASS / 02 (As Per ISO : 8655 - 6 : 2002)

Traceability:

Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.
The following standards / Equipment have been used.

1. Balance With Sensor Cal. Certificate No. MTL / TH / DTM / R01 / 09 - 22 (MTL, Barrackpore) (Cal. Date: 16/09/22, Due Date: 16/09/23)
2. Cal. Certificate No. NC-210 (NSTAR, Ahmedabad) (Cal. Date: 16/05/22, Due Date: 15/05/25)

Remarks:

Medical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1	Volume	5	5.1002	5.1206	0.1206	0.039
2	5 µl to 50 µl	25	25.1586	25.2592	0.2592	0.39
3		50	50.1809	50.3816	0.3816	0.39

i) Cubical Expansion co-efficient of pipette material taken as $10^{-5} \mu\text{l} / ^\circ\text{K}$.

- ii) This result has an expanded uncertainty with a coverage factor $k=2$ at approximately 95% confidence level.
- iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration

N.S. - Not Specified

N.A. - Not Applicable

Opinions and Interpretations

Calibrated	✓	Accepted / Valid for use
Intended Use		Rejected / Out of use

Calibration Engineer

Measure Techno Lab

Kolkata
K. Barat

Calibration Engineer

Checked / Approved by:

Quality & Technical Manager

S. Pandey

Form No. - MTL/22/2006

Issue No. : 2 Issue Date : 10.11.06

Rev. No. : 04 Rev. Date : 01.04.22