



# Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area  
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:  
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :  
www.multitechtechnofab.com



CC3225

## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/12197

ULR :  
CC322523000001525F

Calibration Issued on: 23-02-2023

Job / Identification No. : MTE/CAL/JOB/14832

Date of Calibration : 15-02-2023

Item Receipt Date: 09-02-2023

Recommended Date of Next Calibration: 15-02-2024

Calibrated For

OXFORD HOSPITAL (P) LTD.  
Hospital : 305, Lajpat Nagar, Bhagwan Mahavir Marg, Near  
Nakodar Chowk, Jalandhar - 144001 (Pb).

Reference : Order thru Telephonic  
P. O. No. :

Discipline

Mechanical

Group

Volume

Certificate for : Micropipette

Item Location : Lab

Make / Model

Serial / I.D. No.

Inst. Range

Least Count

Unit

Item Condition

NA / NA

NA / NA

1000

0.0

$\mu$ l

Working

Environmental Condition

22.30  $\pm$  0.5  $^{\circ}$ C

52  $\pm$  3% RH

Atm. Pr in (hPa) 980

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard

Make / Model

Certificate No.

Calibration Agency

Valid up to

Semi Micro Balance

DANWER / DWA265

BBT/412/JAN/23

BLUEBOX TECHNO

24-01-2024

Calibration Procedure

Lab Procedure : MTE-2/M/2.9

STD Procedure : ISO 8655-6

RESULTS: In-Lab Calibration

All Readings are in  $\mu$ l, Specified Otherwise.

U.U.C. READING

Volume @ 27 $^{\circ}$ (ml)  
Chanel 1

UNCERTAINTY<sup>^</sup> in (%)

1000

1.00

0.06

**<sup>^</sup>The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.**

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.

Calibrated By Ankur Chaturvedi  
Technical Executive

For and on behalf of  
**Multitech Technofab Enterprises**

Laboratory Head



# Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area  
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:  
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :  
www.multitechtechnofab.com

**MULTI TECH**  
Technofab Enterprises

## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/12197

ULR :  
CC322523000001525F

Calibration Issued on: 23-02-2023

4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

\*\*\*\*\* END OF REPORT \*\*\*\*\*

For and on behalf of  
**Multitech Technofab Enterprises**

Calibrated By Ankur Chaturvedi  
Technical Executive

Laboratory Head