



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name **MEDICARE DIAGNOSTIC CENTER**
Constituent Group **Chemistry II**
PT Item **Lyophilized human serum based**

Lab No **15263**
Date of Result Entered : **20/04/2023**
Date of Report Published : **06/05/2023**

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	295	148.69	8.33	12.39	139 mg/dL	-0.78	1.44
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	913	39.41	13.59	5.35	52.1 mg/dL	2.37	0.35
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	232	2.81	11.25	0.32	2.77 mg/dL	-0.13	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	235	3.21	14.02	0.45	3.21 mg/dL	0.00	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	265	5.74	8.27	0.48	6.5 g/dL	1.60	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	253	3.33	9.36	0.31	4.4 g/dL	3.45	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	252	6.08	12.80	0.78	4.85 mg/dL	-1.58	0.10
8	CHOLESTROL II	CHOD-PAP II	Biosystems	270	121.38	8.81	10.70	127 mg/dL	0.53	1.30
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	240	109.44	11.69	12.79	98.5 mg/dL	-0.86	1.65

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
MEDICARE DIAGNOSTIC CENTER
12 PVL COMPLEX, KANJIKOIL ROAD CHITTODE,
ERODE
TAMILNADU 638102

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****