
Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company





GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B-00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer	M/S. Medicare Diagnostics Centre,	Certificate No:	AMS/LAB/289/2430
Name &	#12, PVL Complex, Kanji Koil Road,	Issuing Date:	25/09/2022
Address:	Chitode, Erode-638102	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Microscope	Id. No:	MCDC/LAB/MS/001
Make:	Deep vision	Department:	Laboratory
Model:	Dv-40	Calibrated On:	24/09/2022
Sl. No:	14842	Calibration Due	1
Туре:	Binocular Opto Sys.	Suggested:	23/09/2023
Range & Least Count:	war.	DUC Condition on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Digital Lux Meter	HTC/LX - 103	PC/01/22/ATL/0265-02	30/01/2023
3	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	238	mA	300	Normal
2	Frequency	50.0	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	231.5	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	238.4	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	8.8	Volts	< 7	Normal

^{*}Electrical Safety: Please check the attached log sheet

Test Results: Light Intensity LUX:

SL.No.	Knob Position	DUC Reading (LUX)	Distance (mm)	
1	Low	800	30	
2	Medium	5800	30	
3	High	8800	30	

Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

Approved by:

(Technical Manager)